

To whom it may concern

Manufacturer Declaration - Safety in flight -

Löwenstein Medical Technology GmbH + Co. KG
Kronsaalsweg 40 ▪ 22525 Hamburg ▪ Germany

We hereby confirm that the following ventilation devices are therapeutic medical equipment:

prisma VENT30 (WM110TD)
prisma VENT30-C (WM110TD)
prisma VENT40 (WM110TD)
prisma VENT50 (WM120TD)
prisma VENT50-C (WM120TD)
prismaAQUA (WM100TH)

The devices are suitable for transport and use in aircrafts and comply with the requirements in accordance with **RTCA/DO-160G Section 21, Category M**.

i. V. Dr. Christoph Lemke
CQO, Director Quality and Regulatory Affairs
Löwenstein Medical Technology GmbH + Co. KG

An alle, die es betrifft

Bestätigung der Herstellerfirma - Sicherheit im Flugzeug -

Löwenstein Medical Technology GmbH + Co. KG
Kronsaalsweg 40 ▪ 22525 Hamburg ▪ Germany

Hiermit bestätigen wir, dass es sich bei den folgenden Beatmungsgeräten um therapeutische Medizinprodukte handelt:

**prisma VENT30 (WM110TD)
prisma VENT30-C (WM110TD)
prisma VENT40 (WM110TD)
prisma VENT50 (WM120TD)
prisma VENT50-C (WM120TD)
prismaAQUA (WM100TH)**

Die Geräte sind für den Transport und die Verwendung in Flugzeugen geeignet und erfüllen die Anforderungen gemäß **RTCA/DO-160G Section 21, Category M.**

with power of attorney

Dr. Christoph Lemke

CQO, Director Quality and Regulatory Affairs

Löwenstein Medical Technology GmbH + Co. KG