

To whom it may concern

Manufacturer Declaration
- Safety in flight -

Löwenstein Medical Technology GmbH + Co. KG
Kronsaalsweg 40 ▪ 22525 Hamburg ▪ Germany

We hereby confirm that the following device is a therapeutic medical equipment:

CARO (LM300CD)

The device is suitable for transport and use in aircrafts and complies with the requirements in accordance with **RTCA/DO-160G Section 21, Category M**. We further confirm that the included battery has a capacity of less than 100Wh.

Julia Lewalter
Regulatory Affairs Manager
Löwenstein Medical Technology GmbH + Co. KG

An alle, die es betrifft

**Bestätigung der Herstellerfirma
- Sicherheit im Flugzeug -**

Löwenstein Medical Technology GmbH + Co. KG
Kronsaalsweg 40 ▪ 22525 Hamburg ▪ Germany

Hiermit bestätigen wir, dass es sich bei dem folgenden Gerät um ein therapeutisches Medizinprodukt handelt:

CARO (LM300CD)

Das Gerät ist für den Transport und die Verwendung in Flugzeugen geeignet und erfüllt die Anforderungen gemäß **RTCA/DO-160G Section 21, Category M**. Wir bestätigen weiterhin, dass die verbaute Batterie eine Kapazität von < 100Wh aufweist.

Julia Lewalter
Regulatory Affairs Manager
Löwenstein Medical Technology GmbH + Co. KG